User's Guide



Temperature Datalogger

Model TH10



CE

Introduction

Congratulations on your purchase of the Temperature Datalogger. This datalogger measures and stores up to 32,000 temperature readings. The user can easily set up the logging rate, high/low alarm and start-mode, and download the stored data by plugging the module into a PC's USB port and running the supplied PC software. Temperature data can then be graphed, printed and exported to other applications. The data logger is supplied with a long-life lithium battery, which can typically allow logging for one year. This meter is shipped fully tested and calibrated and, with proper use, will provide years of reliable service

Features

- Memory for 32,000 readings
- Selectable data sampling rate: 2s, 5s, 10s, 30s, 1m, 5m, 10m, 30m, 1hr, 2hr, 3hr, 6hr, 12hr, 24hr
- Status Indication via Red/Yellow LED and Green LED
- USB Interface for Set-up and Data Download
- User-Programmable Alarm Thresholds for Temperature
- Analysis software used to view graph
- Long battery life

Description

- 1. Protective cover
- 2. USB connector to PC port
- 3. Start button
- 4. Temperature sensor
- 5. Alarm LED (red/yellow)
- 6. Record LED (green)
- 7. Mounting clip



INSTALLING THE GRAPHING SOFTWARE

Install the supplied Windows PC Datalogger Software by placing the supplied program disk in the PC's CD-ROM drive. If the installation program does not automatically open and provide on-screen prompts, open and browse the CD-ROM drive, then find and double-click on the SETUP.EXE file included on the program disk. Follow the on-screen prompts to complete the installation.

Before running the graphing software, please connect the TH10 to the PC and install the USB Driver software as described in the next section.

INSTALLING THE USB DRIVER

Plug the TH10 into an available USB port on the PC. If this is the first time that the TH10 has been connected to the particular port on the PC, the following screen will appear:

Found New Hardware Wiz	ard .	
	Welcome to the Found New Hardware Wizard This wizard helps you install software for:	
	Silabs C8051F320 USB Board	
	If your hardware came with an installation CD or floppy disk, insert it now.	
	What do you want the wizard to do? Install the software automatically (Recommended) Install from a list or specific location (Advanced) 	
	Click Next to continue.	
	<back next=""> Cancel</back>	

Make sure the program disk is still in the PC CD-ROM drive. Select the 'recommended' installation and follow the on-screen prompts to complete the driver installation. If this is a repeat connection of the TH10 to the USB port, the driver installation process will be automatically skipped.

RUNNING THE GRAPHING SOFTWARE

With the TH10 inserted into the PC's USB port, double-click the Temperature Datalogger icon to launch the program. The main software screen will appear with the MENU BAR icons described as below.

Note: When the TH10 is connected to the USB port, the battery inside the device is discharged at a higher rate than normal. To conserve battery life, do not leave the TH10 connected to the USB port for a prolonged period of time

MENU BAR DESCRIPTION

The main menu for the software appears as follows:

A	Extech	Instr	umen	t RT1	0 Gr	aph
File	View	Link	Help			
	1 💻) 🖻	A		*	Q

From left to right the MENU BAR icons represent DATA DOWNLOAD, LOGGER SETUP, FILE OPEN, FILE SAVE-AS, FILE PRINT, VIEW ALL, and ZOOM. In addition, the FILE, VIEW, LINK, and HELP pull-down menus are available here. All of these functions are described in the following paragraphs.

LOGGER SETUP

Click on the computer icon on the menu bar (2nd from left) or select LOGGER SET from the LINK pull-down menu. The Setup window will appear as shown below; descriptions for each field in the Setup window are listed directly below the illustration:

Connected					
Basic Settings]			
Current Time:	09-06-2007 16:27:14	🔍 Manual 🖲 Instant			
Logger Name:	QC Test	Max. 15 characters!			
Sample Points:	32000 Sample Rate(Se	ec.]: 2s 💌			
LED Flash Cycle: @ 10s C 20s C 30s					
Alarm Settings	Alarm Settings				
Temperature: L	Temperature: Low Alarm 0 High Alarm 40 Unit: Celsius 💌				
	Default	Setup Cancel			

- The CURRENT TIME field will automatically synchronize with the PC date and time setting.
- The MANUAL and INSTANT select buttons allow the user to start data logging immediately when the Setup window is exited (INSTANT), or at a later time (MANUAL).
- The LOGGER NAME field allows the user to select a unique name for the TH10.
- The SAMPLE POINTS field instructs the TH10 to take a fixed number of readings.
- The SAMPLE RATE pull-down menu instructs the TH10 to log readings at a specific rate.
- The LED Flash Cycle can be set by the user: longer cycle times help to extend the battery life.
- The ALARM SETTINGS area of the SETUP window allows the user to set high and low temperature limits and disable the flashing of the TH10's LEDs, thereby conserving battery energy.

Once the changes have been made, click on the SETUP button to save them. Press the DEFAULT button to set the TH10 back to factory default conditions. Press the CANCEL button to abort the setup.

- Note: Any stored data will be permanently erased when the user clicks on Setup. To save the data before creating a new setup, click Cancel and then download the data.
- Note: The battery may run out before the TH10 has finished collecting a specified number of sample points. Always ensure that the remaining charge in the battery is sufficient to last the complete duration of your logging exercise. When in doubt, we recommend that you always install a fresh battery before logging critical data.

If INSTANT was selected in the SETUP window, the TH10 begins logging at the instant the SETUP button is pressed. If MANUAL was selected, the TH10 must be manually started by pressing and holding the device's yellow button for approx. 3 seconds, or until the two LEDs on the TH10 flash at the same time.

Logging will continue at the programmed sample rate until the number of readings (sample points) specified in the Setup window is reached.

The green LED flashes once per sample point and the red or yellow LED flashes when user programmed alarm limits are exceeded. For details, see LED status guide.

DATA DOWNLOAD

To transfer the readings stored in the TH10 to the PC:

- A. Connect the TH10 to the same USB port used when the TH10 was initialized.
- B. Open the graphing software program if it is not still running.
- C. Click the Download icon (1st on the left) or select Data Download from the LINK pull-down menu
- D. The Window shown below will appear. Press DOWNLOAD to begin transferring readings

Download	
Press Download t from logger, or Ca	and load data
nom logger, or ca	

If the readings are successfully transferred, the VIEW screen and the SAVE screen (both shown below) will appear. In the SAVE screen, name the file and save it to a convenient location. The data is saved as an .rec file for use in this program only. At a later time, the user can select SAVE-AS and save the data as an Excel, Text, or Bitmap file (explained in next section).

Name	Size	Туре	Date Modified 🔻	-
🕂 QC temp run.rec	63 KB	Logger Docu	9/6/2007 4:34 PM	-
435363.rec	63 KB	Logger Docu	9/6/2007 3:57 PM	
K testttt.rec	17 KB	Logger Docu	9/5/2007 4:35 PM	
🔼 test 5467.rec	1 KB	Logger Docu	9/5/2007 1:20 PM	
🔼 qc test 8-23-07.rec	17 KB	Logger Docu	8/23/2007 2:31 PM	
test45678.rec	1 KB		8/23/2007 9:54 AM	1

In the screen below press VIEW to see the data graphically. If there are no readings currently stored in the TH10, the program will let the user know via a warning screen.

Connected				
Download				
32000 points finished! Press View to show graph, or Cancel to exit.				
Download	View	Cancel		

When VIEW is pressed, the Data Graph window appears

The Data Graph window presents the transferred data in an x-y graph format with Date/Time represented on the horizontal axis and Alarm Limits represented on the vertical axis:



To zoom in and out of data regions, there are several approaches:

- 1) Use the mouse to click and drag a box around any data area to magnify the selected area.
- 2) Click on the zoom magnifier icon tool on the menu bar
- 3) Select VIEW ALL or ZOOM OUT from the VIEW pull-down menu.

In the example graph above, the Temperature is represented by the solid red line (the lines are known as 'traces'). The low and high alarm values for temperature are represented by the dashed red trace.

To customize the graph window select SHOW TRACES, BACKGROUND, GRID LINES, and MASK POINTS from the VIEW pull-down menu. These options provide the following features:

SHOW TRACES: Allows the user to select which data to view as a trace (Temperature & Alarms).

BACKGROUND: Select background color scheme.

GRID LINES: Add or remove grid lines from x and y axes.

MASK POINTS: Places dots on the traces for the actual data points, breaking up the continuous nature of a trace.

FILE OPEN, FILE SAVE-AS

To save the transferred data in a format other than the proprietary .rec format, click the SAVE AS icon from the menu bar (4th from right) or select SAVE AS from the FILE pull-down menu. The data can be saved in the following formats:

TEXT FILE (.txt) EXCEL FILE (.xls) BITMAP FILE (.bmp)

To open an existing data file for viewing on the data graph window, click on the FILE OPEN icon on the menu bar (3rd from left), or select FILE OPEN from the FILE pull-down menu. When prompted, select a .rec file saved earlier.

FILE PRINT

To print a data graph window to a line or network printer, click on the printer icon or select PRINT from the FILE pull-down menu. Note that color graph windows can be accurately printed on color printers.

Note : The TH10 retains data in its internal memory until the user begins a new logging session. Starting a new logging session clears all of the TH10's stored readings; be sure to save previous data before starting a new logging session.

TH10 LED STATUS GUIDE Green LED Record O Alarm O

Red/Yellow LED

LEDs		Meaning	Action
Record	Alarm	No LEDs flashing -No logging in process Or -No battery found Or -Battery completely discharged	Fit battery in place Replace the battery and download the data
Record	Alarm	Green single flash every 10 sec. * -Logging, no alarm**	
	\bigcirc	Green double flash every 10 sec. * -Delayed start	To start, Hold the start button until Green and Yellow LED flash
Record	Alarm	Red single flash every 10 sec. * -Logging, low alarm for RH	
\bigcirc		Red double flash every 10 sec. -Logging, high alarm for RH	
		Red single flash every 60 sec. - Low Battery [†]	Logging, if enabled, will stop automatically. No data will be lost. Replace battery
Record	Alarm	Yellow single flash every 10 sec. * -Logging, low alarm for TEMP	
$ $ \bigcirc	\bigcirc	Yellow double flash every 10 sec. -Logging, high alarm for TEMP	
		Yellow single flash every 60 sec. - TH10 memory full	Download data

00

 \prod

AA

- * To save power, the TH10's LED flashing cycle can be changed to 20s or 30s via the supplied software.
- ** To save power, alarm LEDs for temperature can be disabled via the supplied software.

.

[†] When the battery is low, all operations will be disabled automatically. NOTE: Logging automatically stops when the battery weakens (logged data will be retained). The supplied software is required to restart logging and to download logged data.

Specifications

Temperature	Overall Range	-40 to 70°C (-40 to 158°F)	
	Accuracy (-40 to -10 and +40 to +70°C)	±2°C	
	Accuracy (-10 to +40°C)	±1°C	
	Accuracy (-40 to +14 and 104 to 158°F)	±3.6°F	
	Accuracy (+14 to 104°F)	±1.8°F	
Logging rate	Selectable sampling interval: From 2 seconds up to 24 hours		
Operating temperature	-35 to 80°C (-31 to 176°F)		
Battery type	3.6V Lithium (1/2 AA) (SAFT LS14250, Tadiran TL-5101 or equivalent)		
Battery life	1 year (typ.) depending on logging rate, ar	mbient temperature, & use of Alarm LEDs	
Dimensions/Weight	101x25x23mm (4x1x.9") / 172g (6oz)		

Battery Replacement

It is highly recommended that the battery be replaced every 12 months, or prior to logging critical data.

The TH10 does not lose its stored readings when the battery weakens or when the battery is replaced; the data logging process will however be stopped and cannot be re-started until the battery has been replaced and the logged data has been downloaded to PC.

Use only 3.6V lithium batteries. Before replacing the battery, remove the datalogger from the PC. Follow diagrammatic steps 1 through 4 below.

NOTE: Leaving the RHT10 plugged into the PC USB port longer than necessary will cause some of the battery capacity to be lost.



You, as the end user, are legally bound (**Battery ordinance**) to return all used batteries and accumulators; **disposal in the household garbage is prohibited!**

You can hand over your used batteries / accumulators at collection points in your community or wherever batteries / accumulators are sold!

Disposal: Follow the valid legal stipulations in respect of the disposal of the device at the end of its lifecycle

Copyright © 2007 Extech Instruments Corporation

All rights reserved including the right of reproduction in whole or in part in any form.